

**INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE  
2A & 2B, Raja S.C.Mullick Road, Jadavpur, Kolkata-700032**

**Corrigendum**

**TENDER NOTICE NO: OC/JG/SB/S1/OC-52/2013/14/28 dated 01.07.2014**

**Sub: Tender for HPLC System**

<i>In the Original Tender Notice</i>	<i>After the Pre-bid Meeting</i>
<p><b>Pump:</b></p> <p>Pulsation: 0.5 MPa or better</p> <p><b>Photodiode Array detector:</b></p> <p>Wave length accuracy check facility should be available with Mercury lamp.</p> <p>Flow cell must be temperature controlled from ambient to 50°C or higher.</p>	<p><b>Pump:</b></p> <p>Pulsation: 0.1 MPa or better</p> <p>Should have liquid leakage sensor for safety reason.</p> <p><b>Photodiode Array detector:</b></p> <p>Wave length accuracy check facility should be available with Mercury lamp or with any other suitable technology, which clearly mention that the wave length accuracy is checked automatically.</p> <p>Should have peak purity analysis feature.</p> <p>Detector should have temperature controlled flow cell (preferably ambient to 50 °C) to prevent deterioration of the spectral resolution and to improve the baseline stability, signal reliability and retention time reproducibility of the analyte.</p>
<p><b>GENERAL INSTURCTIONS:</b></p> <p>10. The tender should accompany a detailed technical compliance chart.</p>	<p><b>GENERAL INSTURCTIONS:</b></p> <p>10. The tender should accompany a detailed technical compliance chart with part no of each component.</p>

<b>Last date and time of submitting tender</b>	<b>August 01, 2014 at 12-30 P.M.</b> □
<b>Date and time of Technical Bid opening</b>	<b>August 01, 2014 at 3.00 P.M. J.C. Bose Hall.</b>

Date:18.07.2014

Registrar