

**Indian Association For The Cultivation of Science**  
**2A &2B Raja S.C. Mullick Road, Jadavpur, Kolkata-700032**

**Corrigendum**

**Tender notice number :SPEC/PCS/14/34**

On the basis of discussion in pre-bid meeting held on 01/09/2014, following amendments may be noted in the technical specifications:

<i>In the Original Tender Notice</i>	<i>After the Pre-bid Meeting</i>
<b>Dual Beam UV-Vis spectrometer required with central PC control using window based user interface</b>	<b>Symmetrical double beam or Dual beam UV-Vis spectrometer required with central PC control using window based user interface</b>
<b>Scan Speed should be very high &gt;15000 nm/minute</b>	<b>Scan Speed should be 6000 nm/minute or more</b>

**Other Terms and Conditions remain unchanged.**

Registrar  
02/09/2014