

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE
2A & 2B Raja S. C. Mullick Road
Jadavpur, Kolkata 700 032

Quotation Notice No.: IACS/BC/PKD/14/51

dated 05.11.2014

Sub: Purchase of MICROPLATE ELISA READER

Through this notice, quotations for a Microplate Elisa Reader are invited. The technical specifications are mentioned below.

TECHNICAL SPECIFICATION FOR MICROPLATE ELISA READER

Parameters	Specifications
Detection mode	Absorbance
Lamp source	Preferably Tungsten halogen
Measurement range	0-3.0 OD or higher
Reproducibility/precision	Preferably <0.5% at 2.5 OD
Accuracy	Preferably <1.0% at 2.5 OD
Wavelength selection	Filters
Filter wheel capacity	5 positions or higher
Standard filters	405, 450, 490, 570, 630 nm and others
Wavelength range	Preferably 400 nm - 750 nm
Plate types	Flat and round 96-well plates and other types well plates
Data output	Export to Printer with the additional option to computer
Reading speed	30 seconds
Resolution	0.001 OD
Shaking	Three steps linear shaking is preferred
Preferable Computer interface	USB

Quotation must reach the office of the **Biological Chemistry, Indian Association for the Cultivation of Science, 2A & 2B Raja S. C. Mullick Road, Jadavpur, Kolkata 700032** on or before the scheduled date & time specified below.

Quotation Notice Number:	IACS/BC/PKD/14/51 dated 05.11.2014
Last date and time of submitting Quotation	27th November, 2014 at 12.30 P.M.
Date and time of opening tender:	27th November, 2014 at 3.30 P.M.
Place of Opening Tender:	at IACS
Contact:	bcpkd@iacs.res.in

General

1. Incomplete & conditional quotations and quotations received after the due date will be summarily rejected without assigning any reasons thereof.
2. The outer cover of the envelop should be super scribed as "**Quotation for purchase of "MICROPLATE ELISA READER"**".
3. The quotation submitted shall remain valid at least for three months from the date of opening the quotation.
4. The rate should be inclusive of all taxes, transportation etc. Nothing extra will be paid in addition to the quoted rate. Price should quoted in Indian Rupee.
5. For any clarification regarding specifications, etc. please send your queries to "**Prof. Prasanta K Das**" at <bcpkd@iacs.res.in>.
6. In the event of date being declared a closed holiday for purchaser's office, the due date for submission of quotation and its opening will be the following working day at the appointed time.
7. **Acceptance of Quotation:** IACS authority does not bind itself to accept the lowest quotation and reserves the right to reject any or all the quotations received without assigning any reason thereof in the interest of IACS.
8. Payment: 100% against delivery and successful installation.
9. Warranty: Minimum 3 years with annual maintenance contract (free of additional charge).
10. Installation/Demonstration /Application Training at site: Free of cost by the supplier.
11. User list: The supplier should provide a list of current users (with contact no, Email address etc) for the quoted item in India.
12. Service facility: Supplier should mention their details of service setup and Manpower in Kolkata who are responsible for after sales support. Response time should be within 24 hours.
13. Validity of quotations beyond three months from the date of opening of the quotation shall be by mutual consent.

REGISTRAR