

NDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE

2A& 2B, Raja S.C. Mullick Road, Jadavpur, Kolkata-700032, India

Corrigendum

Tender Notice No: IACS /Mat. Sc. / N. Pradhan / 2015-16/ 50 Dtd. 09/02/2016

Subject: Procurement of Inert Gas Glove Box

In the original Tender Notice	After corrections
Dust filter 0.3 micron, class H13	Dust filter 0.3 micron, class H13 (or better)
Large Antechamber- diameter minimum 250 mm and minimum 500 mm.	Large Antechamber- diameter minimum 250 mm and Length minimum 500 mm.