

**INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE
Jadavpur, Kolkata - 700032**

Corrigendum-I

To

Tender Notice No: IACS/BC/PKD/DST/2012-13 dated 23.05.2012

[after the Pre-bid Meeting (dated 11th June, 2012)]

Subject: Dynamic Rotational & Oscillatory Research Grade Rheology Instrument with Peltier temperature Controller

On the basis of the Pre-bid meeting held on 11th June, 2012 the following modifications in the tender notice may please be noted:

<i>In the Original Tender Notice</i>	<i>After the Pre-bid Meeting</i>
Point Number 18 in technical specification: Peltier system: The system should be offered with built in Peltier System with a Temperature Range from -40 °C to 200 °C with Auto-recognition	Peltier system: The system should be offered with built in Peltier System with a Temperature Range from -40 °C to 200 °C with Auto-recognition for parallel plate geometry
Point Number 20 in technical specification: Measuring Systems: Concentric Cylinder	Measuring Systems: Concentric Cylinder (25-30 mm)

Other Terms & Conditions remain unchanged.

Registrar