

**INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE**  
**Jadavpur, Kolkata – 700032**

Purchase of Dynamic Rotational & Oscillatory Research Grade Rheology Instrument with Peltier temperature Controller (Tender No.: DST/OC/PD/13/30 dated 30<sup>th</sup> July, 2013)

**Corrigendum-I** (dated 23<sup>rd</sup> Aug, 2013) after the Pre-bid Meeting

**Subject: Dynamic Rotational & Oscillatory Research Grade Rheology Instrument with Peltier temperature Controller**

On the basis of the Pre-bid meeting held on 14<sup>th</sup> August, 2013 the following modifications in the tender notice may please be noted:

<i>In the Original Tender Notice</i>	<i>After the Pre-bid Meeting</i>
1. Motor technology Motor inertia was not mentioned.	If applicable, motor inertia may be mentioned.
2. Motor technology, point no. 4. Maximum torque 10-200 mNm	It should amended as Maximum torque 200 mNm; Minimum torque 10 nNm
3. Point no. 17. Measuring systems: Cone plate (1°, 2° of 25 mm and higher with 50 mm plate), parallel plate (50 mm or higher).	It should be amended as Measuring systems: Cone plate (1°, 2° of 20-50 mm or higher), parallel plate (20-50 mm or higher).
4. Air compressor	Both indigenous and imported air compressor should be quoted mentioning decibel levels