

### **Corrigendum-I**

**Tender Notice No: IACS/SP/PCS/13/35 dated 18.09.2013**

***Subject: Tender for “FTIR spectrometer with ATR accessory”***

On the basis of Pre-bid meeting held on September 30th, 2013 the following modifications in the tender notice may please be noted:

<i><b>In the Original Tender Notice</b></i>	<i><b>After the Pre-bid Meeting</b></i>
<b>Wavelength Range:</b> 8000-6350 $\text{cm}^{-1}$ or further extended	<b>Wavelength Range:</b> 7800-6350 $\text{cm}^{-1}$ or further extended
<b>Signal-to-Noise:</b> 15,000:1 peak-peak 5 seconds scan; 50,000:1 peak-peak 1 minute scan	<b>Signal-to-Noise:</b> More than 10,000:1 peak-peak 5 seconds scan; 45,000:1 peak-peak 1 minute scan
<b>Detector:</b> Temperature stabilized DTGS detector, preferred	<b>Detector:</b> Temperature stabilized DTGS/DLATGS detector, preferred
	<b>New points added No Mechanical or Air Bearing should be used in the spectrometer</b>

The last date of submission and opening of tender is amended as follows:

- 1) Last date & time of submission of tender: **9th October 2013 till 11.00 A.M.**
- 2) Date & time of opening of tender: **9th October 2013 till 3.00 P.M.**

**Other Terms and Conditions remain unchanged.**

**Registrar**  
1.10.2013