

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE

JADAVPUR, KOLKATA-700032

Global Tender No: IACS/SSP/AJP/DIT/10/12-13

Addendum to technical specifications of **“Glove Box fitted with Vacuum Deposition Chamber, Spin Coating Unit, along with a Solar Simulator”**

1. The glove-boxes should be equipped with PLC controlled oxygen and moisture sensors separately for both of the glove-boxes with suitable displays.
2. There should be a purifier with a flow-rate of 80 m³/liter or more for oxygen removal (capacity 30 liter or more) and moisture removal (1.2 kg or more) with fully automatic PLC controller and display separately for each glove-box.

All other technical specifications and terms & conditions of the above-referred tender (including the last date of submission) shall remain unchanged.