

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE

2A & 2B, Raja S.C.Mullick Road

Jadavpur, Kolkata - 700032.

Corrigendum

For the Tender No.: MS/AP/13/31 dated 16.08.2013

On the basis of the Pre-bid meeting held on 18.09.2013 at IACS the following modifications highlighted as underlined text in the respective fields may please be noted:

1. TECHNICAL BID:

The required specifications are as follows:

- A. **Monochromator** with usable wavelength range between 200 nm to 2.4 micron dependent on installed gratings. (focal length at least 150 mm)
- B. **Light Source** Tunable output power adjustable from 200-500 W with Xe or dual Light Source.
- C. **Lock in amplifier/meter** with control unit detectors with the preamplifier and optical chopper.
- D. **Si Detector** with working range 200 - 1100 nm. (Up gradation optional)
- E. The QE measurement system should be compatible for photovoltaic measurement is preferred. Interface facilities with the computer and Electrometers are desirable.
- F. Warranty for minimum 2 year after installation.
- G. Installation, personnel training and test measurements.

2. BID SECURITY:

2) Bid Security (EMD): An account payee Demand Draft of Rs. 1.0 Lakh in the name of "Indian Association for the Cultivation of Science" is to be furnished by the bidders as Bid Security or Earnest Money Deposit (EMD).

Registrar