

Invitation for Tender for Supply of a Continuum Xe Lamp Light Source
(Tender Reference No. IACS/SKD/MS/01/2012-13)

Indian Association for the Cultivation of Science invites Tenders for a Continuum.Xe Lamp Light Source. The bids are to be submitted in two bid pattern i.e. Technical Bid and Price Bid in two separate sealed covers distinctly marked accordingly and both to be put inside another envelope, which should be sealed and **superscribed with Xe Lamp Light, Tender reference no., due date and time of opening.** The two bids, i.e. Technical Bid and Price Bid should be identical in all respect that the technical Bid should have blank spaces at the places where prices have been stated in the Price Bid.

Interested vendors may please send their quotations to:

Prof . Subodh Kumar De
Department of Materials Science
Indian Association for the Cultivation of Science
2A, 2B Raja S. C Mullick Road
Jadavpur
Kolkata 700 032

Tender Reference	IACS/SKD/MS/01/2012-13
Last date and time of submitting tender	August 31, 2012 at 1:00 pm
Date and time of opening tender	August 31, 2012 at 3:00 pm
Place of opening tender	I. A. C. S Kolkata
Contact	Email: msskd@iacs.res.in Tel: +91-33-2473 4971 (Extn. 1206)

If holiday is declared on the scheduled date of opening, the quotation will be opened on the next working day at 3:00 P.M. The technical bids will be opened first to evaluate technical specifications of the equipment and there after the price bids of only the technically qualified companies will be opened.

A) TECHNICAL BID:

Technical Specifications for the Xenon Light Source power supply

Lamp : Xenon
Power : 150W
Light flux: 3000 lm
Light intensity: 300 cd
Average life cycle: More than 1000 hrs
Operational Spectral Range: 190 nm-1100 nm
Exit aperture diameter: 60mm
Power Supply: 230 + 10% Vac, 50 Hz

B) FINANCIAL BID:

The financial bid should be super scribed as “**Finance Bid**” on the outer cover of the envelope.

C) GENERAL INSTURCTIONS:

i) Submitting Tender: The tender should not contain any material (eg. brochures, company profiles etc.) other than the two above-mentioned covers containing specifically the items requested. the vendors are allowed to deviate from the specifications given below only when such deviation are demonstrated to be technically superior. Additional features in the quoted items which are better than those in the indent- may be highlighted.

The technical bid should not have any mention of pricing. The commercial bid should have the pricing for each option separately. Prices should be inclusive all charges (taxes,/duties/as applicable, shipping charges, delivery at site, installation etc.) with clearly indicated breakups. Fax, e-mail Tender will not be accepted. Duplicate bid document must be submitted in separate closed cover.

ii) Tender updates: Prospective bidders please refer to our website <http://www.iacs.res.in> for any changes which may appear from time to time.

Important information

1. I.A.C.S., Kolkata may accept or reject any or all the bids in part or in full without assigning any reason. The association at its discretion may change the quantity/upgrade the criteria/ drop nay item or part thereof at any time before placing the Purchase Order.

2. Tender submitted shall remain valid at least for three months from the date of opening the tender. Validity beyond three months from the date of opening of the tender shall be by mutual consent.

3 Incomplete tenders will be summarily rejected.

4. In case of any dispute, the decision of the Institute authority shall be final and binding on the bidders.

5. For any query pertaining to this bid document correspondence may be addressed to Professor Subodh Kumar De, at the address mentioned above.

Registrar